the bioavailability of drug around 25% of the nominal dose. If the same ratio applies for young children it can be assessed that the bioavailability of budesonide inhaled from a Nebuhaler will be around $55/90 \times 25\% = 15\%$ of the nominal dose, that is 40% lower than for schoolchildren. When such assessments are made it should be remembered that young children are more prone to inhale through the nose than older children.³ This may affect the clinical effect as well as the systemic effect of the inhaler but probably to the same extent for all the inhalers studied.

The total dead space added to the inhalers in the study was 45 ml. Although this volume was the same for all devices, it caused a higher percentage increase in volume for the low volume spacers than for the Nebuhaler. However, the volume inhaled by the children during 30 seconds by far exceeded this volume plus the volume of the spacer. Therefore, we do not believe that this increase in volume has influenced the results to any clinically important extent. The rather high minute ventilation probably also explains why spacer volume did not influence the dose to the patient. In an earlier study we found that attachment of a filter did not significantly affect the breathing pattern of the children. So it is most unlikely that the filter caused the high minute ventilation in the present study.

The information derived in our study applies to children studied awake. The results might be different if the treatment had been given during sleep as it is not known if the minute ventilation would increase to the same extent under such circumstances.

The finding that the Nebuhaler delivered more drug to the patient than the Aerochamber is in good agreement with the results in a recent *in vitro* study, which found that drug delivery was greatest from the Nebuhaler at tidal volumes similar to those measured in the present study.⁴ Only at six breaths of tidal volumes around 25 ml did the Aerochamber enhance drug delivery as compared with the Nebuhaler.⁴ The clinical effect of the two inhalers has not been compared. However, in schoolchildren the Aerochamber resulted in less bronchodilatation than the Inspirease spacer,^{5,6} probably as a result of the lower output from the Aerochamber.

Apart from measuring the mean dose deposited on the filter it is also important to evaluate the dose range for each inhaler as this gives some idea about the reproducibility of the dose delivered to the patient. A 3.5-fold variation in dose was seen for the Aerochamber whereas a 2-fold variation was observed for the two other inhalers, indicating that these would deliver a more consistent dose to the patient. However, further studies on the intraindividual variation in dose are needed before firm conclusions can be made. In the present study we evaluated patients who had never used a spacer device before. Furthermore, it is possible that training will reduce the variation in dose delivery.

The results of comparative studies in new users of inhalation devices may be influenced

by the experience of the investigator in the use of the various inhalers. In our clinic about 50% of young children normally use an Aerochamber, 45% a Nebuhaler, and about 5% a Babyspacer. So a negative effect in disfavour of the Babyspacer results cannot be excluded. However, differences in experience could not explain the poorer results after Aerochamber inhalations. Differences in design, plastic material, and valve system may all have caused a higher retainment of drug in the Aerochamber. The higher variation in dose is most likely to be due to differences in valve system.

With these reservations we think that the study provides some useful information about inhalation treatment in young children with asthma. The larger variation in dose and the lower dose delivered from the Aerochamber suggest that this device is not (as we believed before the study) preferable to the two other spacers in the treatment of very young children (older than one year). Furthermore, the finding that children tend to increase their minute ventilation when a facemask is applied to their face, suggests that spacer volume is not so critical as normally believed. This observation means that results from studies that assess the effect of an inhaler by setting a respirator to mimic the tidal breathing of a young child should be interpreted with some caution if this factor is not taken into account.

- 1 Lödrup Carlsen KC, Nikander K, Carlsen KH. How much nebulised budesonide reaches infants and toddlers? *Arch Dis Child* 1992; **67**: 1077-9.
- 2 Pedersen S, Steffensen G, Ekman I. Pharmacokinetics of budesonide in children with asthma. *Eur J Clin Pharmacol* 1987; **31**: 579-82.
- 3 Pedersen S, Østergaard PA. Nasal inhalation as a cause of inefficient pulmonary aerosol inhalation technique in children. *Allergy* 1983; **38**: 191-4.
- 4 Everard ML, Clark AR, Milner AD. Drug delivery from holding chambers with attached facemask. *Arch Dis Child* 1992; **67**: 580-5.
- 5 Crimi N, Palermo F, Cacopardo B, *et al.* Bronchodilator effect of Aerochamber and Inspirease in comparison with metered dose inhaler. *Eur J Respir Dis* 1987; **71**: 153-7.
- 6 Lee H, Evans HE. Evaluation of inhalation aids of metered dose inhalers in asthmatic children. *Chest* 1987; **91**: 366-9.

Commentary

Although at first reading, one might be forgiven for thinking that this is just another study of a few of the potentially infinite combinations of puffers, spacers, facemasks and young children, it does have one important feature that distinguishes it from many other studies: the measurements are clinically relevant, albeit only under the precise conditions which applied during the study.

When one considers the huge effort that has gone into the development and marketing of topical asthma drugs, the trivial investment that has gone into the development of therapeutic devices for preschool children seems almost negligent. Pedersen has been almost alone over the last 12 years in showing, alas often in retrospect, the limitations of the devices which have been 'handed down' to young children.

Each aspect of aerosol administration by metered dose inhaler (MDI), spacer, and facemask needs careful study: the size, shape,

and material construction of the spacer itself in relation to the formulation and jet velocity of the MDI; the resistance, dead space, and disturbance introduced by valves (and even their necessity); the interaction between these 'physical' features and breathing patterns during wakefulness and sleep during aerosol administration; the dose available at the airway opening; the effects of nasal filtration; and finally the pulmonary dose and distribution. Attempts at administration in the less than ideal circumstances surrounding an acute attack of asthma in a 2 year old in a darkened bedroom at 2 am may, of course, render all

of these data largely irrelevant! But such systematic information would at least allow comparisons between devices and improvements in design to be carried out.

Paediatricians have been largely unquestioning in their acceptance of spacer devices. Perhaps we should demand that the regulatory authorities consider a licensing procedure for the devices which we use to administer aerosols rather than simply considering the agents which go into them.

MICHAEL SILVERMAN

*Department of Paediatrics and Neonatal Medicine,
Royal Postgraduate Medical School,
London W12 0NN*