Abstract PS-376 Figure 1

Results 50 patients were enrolled (53 male). Main results are shown in table 1. LUS diagnosis were made based in Lichtenstein and Coppetti standards.

Conclusions Lung ultrasound can differentiate those patients with NNT from other causes of neonatal respiratory distress may be an extremely useful tool for respiratory prognosis and can reduce the number of chest radiographs performed in these patients.

PS-376 NON-INVASIVE VENTILATION IN NEONATES: EFFECT OF NASAL CANNULA SIZE, INSERTION DEPTH AND NARES DIAMETER ON MEAN AIRWAY PRESSURE - AN IN-VITRO STUDY

J Gerdes, E Siveri, S Abbasi. *Pediatrics Division of Neonatology, Pennsylvania Hospital and Children's Hospital of Philadelphia University of Pennsylvania, Philadelphia, USA*

10.1136/archdischild-2014-307384.675

Background Increased neonatal Non-Invasive Ventilation (NIV) use, such as nCPAP and high flow nasal cannula (HFNC) prompted re-evaluation of nasal interface devices – specifically short bi-nasal prongs, and their size and fit relative to the nares. **Objective** To quantify the effect of nasal prong size and insertion depth on generated mean airway pressure (MAP) during NIV.

Design/methods A Dräger traditional nCPAP interface, three sizes of RAM cannulae (Neotech, Valencia, CA, USA) and two Fisher and Paykel (FP) (Auckland, NZ) HFNC cannulae were tested with 8 simulated nares sizes. A simulated nasal airway was connected to an active lung model set at: V_t 8–10 ml, 60 b/m, T_i 0.35–0.40 s. A Dräger Evita XL ventilator delivered 4,5,6,7 and 8 cmH₂O to the RAM and the Dräger nCPAP cannulae and a FP HFNC system delivered 1–6 L/min flows. MAP was measured for open and closed-mouth conditions.

Results MAP decreased progressively as the percent nares occlusion (%Occl) decreased. At one-half insertion and closed-mouth conditions, close fitting prongs with high % Occl yielded MAP's of one-half of set CPAP levels. Delivered MAP's were significantly lower during open-mouth condition using either HFNC or CPAP cannulae. MAP increased with both flowrate and % Occl. A rapid rise in pressure was associated with %

Occl's >74%. Complete nares occlusions generated the highest pressures. Simulated closed-mouth produced 12 ± 7 SD

higher MAP's than open-mouth over all %Occl's at HFNC flows >1 L/min.

Conclusion RAM 'CPAP' does not deliver MAP as effectively as standard nasal prong CPAP. Optimum delivered airway pressure is dependent on appropriate %Occl, insertion depth and NIV settings.

PS-377 WITHDRAWN

PS-378 IMPACT OF AMPLITUDE, FREQUENCY AND I:E RATIO ON TIDAL VOLUME IN HIGH FREQUENCY VENTILATION (VTHF) USING DIFFERENT HYBRID NEONATAL VENTILATORS

C Mendez, MB Milet, A Maturana. *Neonatology, Clinica Alemana, Santiago, Chile*

10.1136/archdischild-2014-307384.676

Background and aims High frequency ventilation (HFV) is frequently used in both term and preterm newborns. Hybrid neonatal ventilators capable of both HFV and conventional ventilation are currently available with different mechanisms for

Abstract PS-378 Figure 1

Abstract PS-378 Figure 2

generating the required high frequency oscillation. In HFV, carbon dioxide elimination is directly correlated with VThf. The aim of the current study was to describe the impact of changes in amplitude, frequency and inspiratory-expiratory ratio (I:E) on VThf using different hybrid neonatal ventilators.

Methods Three different hybrid neonatal ventilators were tested: Dräger Babylog® VN500, Acutronic Fabian and SLE 5000. Each ventilator was calibrated using a new disposable ventilator circuit with a Dräger pneumatic test lung leaving the humidifier dry to standardise all measurements. The VThf measured by the device was recorded after the system stabilised for 3 min. Each experiment was done using 21% oxygen, 8 cm H₂O of pressure and repeated 3 times with negligible differences between measurements.

Results In this setting, there is an expected inverse relationship between frequency and achieved VThf in all devices when using maximum amplitude. All three ventilators were more efficient at lower frequencies achieving higher VThf. In the two ventilators that allow I:E modification, the 1:1 ratio achieves higher VThf than 1:2 ratio. This becomes more evident at lower frequencies and higher amplitudes.

Conclusions In tested ventilators the highest impact on VThf was achieved with frequency reduction. When modifiable, an I:E ratio of 1:1 optimises VThf at higher amplitudes. Clinicians need to be aware of these characteristics for optimising ventilation strategy.

PS-379 THE OPTIMAL PREDICTORS OF READINESS FOR EXTUBATION IN LOW BIRTH WEIGHT INFANTS

W Janjindamaj, S Pasee. *Pediatrics, Prince of Songkla University, Songkhla, Thailand*

10.1136/archdischild-2014-307384.677

Background and aims Reintubation, following an unsuccessful extubation from mechanical ventilation is traumatic to the infant and the family. However 20–40% of infants fail extubation and reintubation. The aim of the study was to determine the optimal predictors of readiness for extubation in low birth weight infants during endotracheal tube-continuous positive airway pressure (ETT-CPAP) for three minutes.

Methods A prospective cohort study was undertaken in 51 mechanically ventilated infants who were considered to be ready for extubation. The infants were changed to ETT-CPAP for a 3-minute spontaneous breathing test (SBT) before extubated. Infants were divided into two groups based upon whether they failed or passed the extubation attempt. Extubation failure was defined as reintubation within 72 h of extubation.

Results Forty-five of 51 infants (88%) were successfully extubated. Out of the 51 infants only one infant failed the SBT. The three predictors of extubation success that included the SBT, ratio of minute ventilation during ETT-CPAP to mechanical ventilation and ratio of respiratory frequency during ETT-CPAP to mechanical ventilation were not significantly different. Using synchronised nasal intermittent positive pressure ventilation after extubation in the failed extubation group was significantly higher than the successful extubation group (66.7% vs 15.7%, $p = 0.02$).

Conclusion The SBT and minute ventilation ratio in low birth weight infants were not optimal predictors of readiness for extubation. However, a further prospective study in this field with a larger number of subjects and a proper indication for extubation should be considered.

Abstract PS-379 Table 1 Clinical characteristic of the infants

Clinical characteristic	Failed extubation n = 6	Successful extubation n = 45	P value
Gestational age (week, SD)	29 (2.4)	30.5 (3.6)	0.31
	1,222.5	1,205	
Birth weight (g, IQR)	(1,146, 1,269)	(1,020, 1,775)	0.59
Male n, (%)	3 (50)	22 (48.9)	1.00
Vaginal delivery n, (%)	1 (16.7)	20 (44.4)	0.38
Weight at extubation mean, g (SD)	1,345 (220.1)	1,549.1 ± 496	0.99
Apgar score at 5 min (median, IQR)	8 (5.8, 8)	8 (6, 8)	0.48
Duration of ventilation, mean (day, IQR)	22.5 (12, 47)	12 (5, 21)	0.08
Methylxanthine used, n (%)	4 (66.7)	22 (48.9)	0.67

Abstract PS-379 Table 2 Pulmonary function test and predictive values during on continuous positive airway pressure (ETT CPAP)

Indicators	Failed extubation n = 6	Successful extubation n = 45	P value
ETT CPAP (cmH ₂ O) (median, IQR)	3 (3, 3)	3 (3, 3)	0.63
FiO ₂	0.3 (0.21, 0.3)	0.3 (0.21, 0.3)	0.71
Pass Spontaneous breathing test, n (%)	6 (100)	44 (97.8)	1.00
Expiratory tidal volume (ml/kg.) (median, IQR)			
Expiratory minute ventilation (l/min/kg.) (mean, SD)	5.1 (3.3, 6.7)	6.2 (4.6, 7.3)	0.36
Frequency (/min) (median, IQR)	0.3 (0.2)	0.4 (0.1)	0.51
Ratio V _{E TOT} ≥ 0.8	81 (76.2, 83.5)	68 (60.81)	0.23
Ratio V _{E TOT} ≥ 0.5	4 (66.7)	41 (91.1)	0.14
Ratio fr ≤ 1.5	2 (33.3)	25 (55.6)	0.40
	1 (16.7)	2 (4.4)	0.32

Ratio V_{E TOT}, ratio of mean minute ventilation between breathing on ET CPAP and mechanical ventilation;
Ratio fr, ratio of mean respiratory rate between breathing on ET CPAP and mechanical ventilation.